

Защитная трёхслойная маска для лица

Трёхслойная защитная маска необходима для посещения общественных мест и при контакте с людьми. Обеспечивает надёжный уровень фильтрации от бактерий и вирусов, не затрудняя дыхания.

HOBMIN

Предназначена только для индивидуального одноразового использования.

Изготовлена из синтетического нетканого материала. Первый и третий слои маски изготовлены из спанбонда, средний (фильтр) из мельтблауна.

В области переносицы для лучшего прилегания к лицу предусмотрен носовой фиксатор.


Тип крепления: на резинках. Размер: 175х95 мм.





Use melthod





Fix the nose bridge to the nose to prevent dirty air from entering. Please pull the mask to the lower jaw to immediately present a three-dimensional breathing space.

Sandard - GB/732610-2016

and non-medical, non st erile)

Protecting People's Health Globally SOFT-BREATHABLE-SKIN-FRIENDLY

Company Name: Shenzhen KeYueDuo Intelligent Electronics Co., Ltd. Address: F-5, Bldg.B, HeDong Area, HangCheng Ind. Area, TaoYuan Community, XiXiang Sub-dist., Bao'an Dist., Shenzhen, Guangdong, China



2-Layer Non Woven Fabric+1-Layler Melt Blown Fabric BFE≥95%

Disposable Mask 3-Layer Ear-loop Type

•Effective protection •Stretchable elastic ear loops •Comfortable material

Made in China

Model: 2020-K11 CE EN 149:2001+A1:2009

50PCS









Certificate

ICR Polska/CE/V/ 00RE210

Name and address of certificate owner:	Shenzhen KeYueDuo Intelligent Electronics Co., Ltd.
Name and address of manufacturer:	F-5, Bldg.B, HeDong Area, HangCheng Ind. Area, TaoYuan Community, XiXiang Sub-dist., Bao'an Dist., Shenzhen, Guangdong, China
Product name:	Mask
Product types:	2020-K11
Product trademark:	N/A
• This certificate confirms that the product meets the requirements of the following standard	

- This certificate confirms that the product meets the requirements of the following standard
 The conformity was demonstrated based on following standard(s) listed by European Commission
- as harmonized with Regulation 2016/425

EN 149:2001+A1:2009

The certification has been carried out in accordance with individual rules and conditions agreed with the applicant. Evaluation has been carried out in accordance with:

Test report(s) No.:	NCTC20200302151
Test conducted by:	Shenzhen Zhongcheng Testing Technology Co., Ltd.
Certificate issue date:	02.04.2020
Expiration date:	01.04.2025

NOTES:

- This certificate refers to the above mentioned product and its conformity in regards of above mentioned standard(s) was proven on test sample
- This certificate does not imply meeting all essential requirements, assessment of the series-production or any other restricted Notified Bodies conformity assessment procedure appropriate for the product
- This certificate holder shall use this certificate in connection to declaration of conformity and technical data relevant for the product the certificate was issued

Certificate modifications: n/a



B

Director: Rafał Kalinowski

Warsaw, 02.04.2020

ICR Polska Co. Ltd. Plac Przymierza 6 03-944 Warsaw, POLAND www.icrpolska.com e-mail: icrpolska@icrqa.com

